

Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N°
R60/1991-NL-95.05
Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI IJkwezen B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: M. Charité

Applicant

Name: Revere Transducers Europe BV
Address: Ramshoorn 7
4824 AG Breda
The Netherlands

Manufacturer of the certified pattern

Name: Revere Transducers Europe BV
Address: Ramshoorn 7
4824 AG Breda
The Netherlands

Identification of the certified pattern

Type : HPS
Fraction : $P_1 = 0.7$

Maximum Capacity (E_{max})	6, 12, 30 and 60 kg		
Accuracy Class	C1	C2	C3
Maximum number of load cell intervals (n)	1000	2000	3000
Minimum load cell verification interval (V_{min})	$E_{max} / 5000$	$E_{max} / 10000$	$E_{max} / 12000$

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht (NL)
Telephone +31 78 33 23 32
Telefax +31 78 33 23 09

Nederlands Meetinstituut N.V. (Registered at the Chamber of Commerce Delft number 28701)


Subsidiary companies:
NMI Certin B.V. (33418)
NMI Van Swinden Laboratorium B.V. (28703)
NMI IJkwezen B.V. (28700)
NMI Test- en Adviescentrum (TAC) B.V. (28702)

This certificate is issued under the provision that Nederlands Meetinstituut N.V. nor its subsidiary companies accept any liability.

Reproduction of the complete certificate is allowed. Parts of the certificate may only be reproduced after written permission.



QUALIFIED BY STERLAB
Reg. nr. I 029



Nederlands Meetinstituut

OIML Certificate N°

R60/1991-NL-95.05

Page 2 of 2

This certificate attests the conformity of the above-mentioned pattern (represented by the sample identified in the associated test report, the test certificate and the description with number TC2513 and the documentation folder number TC2513-1, with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML):

R60
edition 1991 (E)
for accuracy class C and D

This certificate relates only to the metrological and technical characteristics of the pattern of the Loadcell concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report N° R60/1991-NL-95.05, that includes 38 pages.

The issuing authority
M. Charité

31 January 1995

The OIML member
G.J. Faber

31 January 1995

*
* *

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.