



Member State of OIML  
United Kingdom of Great Britain  
and Northern Ireland

OIML Certificate No  
R60/1991-GB-95.07

## OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name:

**National Weights and Measures Laboratory**

Address:

**Stanton Avenue  
Teddington  
Middlesex  
TW11 0JZ**

Person responsible:

**M D Koch - Head of Type Examination**

Applicant

Name:

**Revere Transducers Europe**

Address:

**Ramshoorn 7  
Postbus 6909  
4802 HX Breda  
The Netherlands**

Manufacturer of the certified pattern .

**Revere Transducers Inc  
14192 Franklin Ave,  
Trustin, LA 92680-7016 USA**

Identification of the certified pattern:

**Load Cell Model No 5222 / 5223  
Further characteristics see page 2**

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R60 edition 1991 (E) for accuracy class C**

OIML Certificate No  
R60/1991-GB-95.07

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report application no 0195, that includes 21 pages.

The issuing authority

NWML

Date: 30 March 1995

The OIML member

S J Bennett

Date: 30 March 1995

*S J Bennett*

Characteristics:

type	Class	max output	input impedance	output impedance	cable type	cable length
5222	C	2 mV/V	700 $\Omega$	700 $\Omega$	4 wire shielded. Shield not connected to load cell	12 m (40 feet). Other options taken into consideration during production.
5223	C	3 mV/V	700 $\Omega$	700 $\Omega$	4 wire shielded. Shield not connected to load cell	As above.

Type	Capacity $E_{Max}$	$v_{min}$	$E_{min}$	$n_{max}$
5222	20000 kg	$E_{Max}/10000$	0	3000
5222	30000 kg	$E_{Max}/10000$	0	3000
5222	45000 kg	$E_{Max}/10000$	0	3000
5223	50000 lb	$E_{Max}/10000$	0	3000
5223	65000 lb	$E_{Max}/10000$	0	3000
5223	100000 lb	$E_{Max}/10000$	0	3000

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.