

Nederlands Meetinstituut

Test certificate

Number **TC2792** revision 3
Project number 503569
Page 1 of 4

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body Number 0122

In accordance with Paragraph 8.1 of the European Standard on Metrological aspects of non-automatic weighing instruments EN 45501:1992/AC:1993 and by application of the OIML International Recommendation R 60 (Edition 2000). The applied error fraction p_i , meant in the paragraph 3.5.4. of the standard is 0.7.

Applicant Vishay Transducers
5a Hatzoran St.
Netanya, 42506
Israel

In respect of A **single point, bending beam load cell**, with strain gauges, tested as a part of a weighing instrument.
Manufacturer : Vishay Transducers or TedeA-Huntleigh
Type : 1022, 1022P and LPS

Characteristics Model 1022:

Maximum capacity (E_{max})	3 kg up to and including 150 kg			
Accuracy class	C			
Maximum number of load cell verification intervals (n_{max})	1000	2000	3000	4000
Ratio of minimum LC verification interval $Y = E_{max} / V_{min}$	3333	6666	10000	12000

Model 1022P and LPS:

Maximum capacity (E_{max})	7 kg up to and including 150 kg			
Accuracy class	C			
Maximum number of load cell verification intervals (n_{max})	1000	2000	3000	4000 *)
Ratio of minimum LC verification interval $Y = E_{max} / V_{min}$	3333	6666	10000	12000 *)

*) Only valid for $E_{max} \geq 30$ kg.


NMI Certin B.V.
Hugo de Grootplein 1, 3314 EG Dordrecht
P.O. Box 394, 3300 AJ Dordrecht, NL
phone +31 78 6332332
fax +31 78 6332309
certin@nmi.nl
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI B.V. (see "Regulation objection and appeal against decisions of NMI B.V.")

NMI Certin B.V., chamber o.c. no. 27.233.418

This document is issued under the provision that no responsibility is accepted and that the applicant gives warranty for each responsibility against third parties.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission.



Nederlands Meetinstituut

Test certificate

Number **TC2792** revision 3
Project number 503569
Page 2 of 4

In the description number TC2792 revision 3 further characteristics are described.

Description and documentation The load cell is described in the description number TC2792 revision 3 and documented in the documentation folder TC2792-1, appertaining to this test certificate.

Remarks Summary of the test involved: see Appendix number TC2792 revision 3
This revision test certificate replaces the earlier version(s), except for its documentation folder.

Dordrecht, 17 May 2005
NMI Certin B.V.


Ing. C. Oosterman
Manager Product Certification



1 General information about the load cell

All properties of the load cell, whether mentioned or not, may not be in conflict with the standard mentioned in the test certificate.

1.1 Essential parts

Description	Drawing number	Rev.	Remarks
Model 1022 Load cell	187.000.00-3	D	Mechanical
Model 1022P Load cell	218.000.00-3	A	Mechanical
Wired sensor	187.200.00-2	B	Electrical
Model 1022P Load cell	218.200.00-2	1	Electrical

Cable:

- The load cell is provided with a 4 or 6-wire system.
Because (no Remote sensing) is used by the 4-wire system that cable length has to be approximate 1 meter. If the load cell is provided with a 6-wire system (=“Remote-sensing”). The cable length is not limited.
- The cable should be a shielded cable, the shield may be connected to the load cell.

1.2 Essential characteristics

Minimum dead load	: 0 kg
Safe overload	: 150 % of E_{max}
Rated Output	: 2 mV/V \pm 0.2 mV/V
Input impedance	: 415 Ω \pm 15 Ω
Output impedance	: 350 Ω \pm 3 Ω
Recommended excitation	: 10 V DC/AC
Excitation maximum	: 15 V DC/AC
Transducer material	: Anodized Aluminium or Non-Anodized Aluminium
Atmospheric protection	: Adhesive Silicone Rubber

1.3 Essential shapes

The load cell is built according to the drawings:

- Model 1022 Load cell, drawing number 187.000.00-3;
- Model 1022P Load cell, drawing number 218.000.00-3

The data plate is secured against removal by sealing or will be destroyed when removed. The data plate mentions at least the information and markings as described in the OIML R60 document. In the countries where it is mandatory the load cell should bear this test certificate number: TC2792.

Securing:

The connecting cable of the load cell or the junction box is provided with possibility to seal.



Tests performed for this test certificate:

Test	Institute	type, version, remarks
Temperature test and repeatability (20, 40, -10 and 20 °C)	NMi Certin B.V	1022, 3 kg C4 and 30 kg C4 1022P, 7 kg C3 and 30 kg C4
Temperature effect on minimum dead load output (20, 40, -10 and 20 °C)	NMi Certin B.V	1022, 3 kg C4 and 30 kg C4 1022P, 7 kg C3 and 30 kg C4
Creep (20, 40 and -10 °C)	NMi Certin B.V	1022, 3 kg C4 and 30 kg C4 1022P, 7 kg C3 and 30 kg C4
Minimum dead load output return (20, 40 and -10 °C)	NMi Certin B.V	1022, 3 kg C4 and 30 kg C4 1022P, 7 kg C3 and 30 kg C4
Barometric pressure effects at room temperature	NMi Certin B.V	1022, 3 kg C4 1022P, 7 kg C4
Damp heat, cyclic: marked CH (or not marked)	NMi Certin B.V.	1022, 3 kg C4 1022P, 7 kg C4