

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMi Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: P.P.M. van Enckevort

Applicant

Name: Tedea-Huntleigh International Ltd.
Address: 5 Hazoran Street
New Industrial Zone Netanya
Israel

Manufacturer of the certified pattern

Name: Tedea-Huntleigh International Ltd.
Address: 5 Hazoran Street
New Industrial Zone Netanya
Israel

Identification of the certified pattern

Type : 1263
Fraction : $P_1 = 0.7$
Temperature range - 10 °C / + 40 °C

Maximum capacity (E_{max})	50, 67, 100, 150, 200, 250, 300, 330, 500 and 635 kg		
Accuracy Class	C		
Maximum number of load cell intervals (n)	1000	2000	3000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	2000	4000	15000

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the test certificate and the description with number TC6092 and the appertaining documentation folder), with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML):

R60
edition 2000 (E)
for accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report:

- No R60/2000-NL-02. 24A, that includes 40 pages;
- No R60/2000-NL-02. 24B, that includes 38 pages.

This revision replaces the earlier versions.

The issuing authority
P.P.M. van Enckevort
Manager Certification Delft

The OIML member
G.J. Faber

21 August 2002

21 August 2002

*
**

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.